

A Double-Blind Trial of the Clinical Effects of Pulsed Electromagnetic Fields in Osteoarthritis

DAVID H. TROCK, ALFRED JAY BOLLET, RICHARD H. DYER, Jr., L. PETER FIELDING, W. KENNETH MINER, and RICHARD MARKOLL

ABSTRACT. *Objective.* Further evaluation of pulsed electromagnetic fields (PEMF), which have been observed to produce numerous biological effects, and have been used to treat delayed union fractures for over a decade.

Methods. In a pilot, double-blind randomized trial, 27 patients with osteoarthritis (OA), primarily of the knee, were treated with PEMF. Treatment consisted of 18 half-hour periods of exposure over about 1 month in a specially designed noncontact, air-coil device. Observations were made on 6 clinical variables at baseline, midpoint of therapy, end of treatment and one month later; 25 patients completed treatment.

Results. An average improvement of 23–61% occurred in the clinical variables observed with active treatment, while 2 to 18% improvement was observed in these variables in placebo treated control patients. No toxicity was observed.

Conclusion. The decreased pain and improved functional performance of treated patients suggests that this configuration of PEMF has potential as an effective method of improving symptoms in patients with OA. This method warrants further clinical investigation. (*J Rheumatol* 1993;20:456–60)

Key Indexing Terms:
OSTEOARTHRITIS

PULSED MAGNETIC FIELDS

Pulsed electromagnetic fields (PEMF) have been widely used in the treatment of delayed union fractures for over a decade, and there is a growing body of literature concerning the biological and clinical effects of such low frequency, nonionizing forms of energy.

Clinical responses have been reported in longstanding nonunion of fractures¹, failed arthrodeses², avascular necrosis of hips in adults^{3,4}, and Legg-Perthes's disease in children⁵. Similar devices are being evaluated in the treatment of osteoporosis. A device to generate PEMF installed in a body brace increased the success rate of lumbar fusions in a double blind control study⁶. There are also reports of augmentation of peripheral nerve regeneration and function and promotion of angiogenesis⁷. Patients with persistent rotator cuff tendinitis, refractory to steroid injection and other conventional measures, showed significant benefit compared

with placebo treated patients⁸. In clinical studies done over 17 years in over 200,000 patients treated safely with PEMF, no toxic effects have been reported⁹.

One of us (RM) developed a unique delivery system of extremely low frequency (ELF) pulsed waves, consisting of a magnetic field generator with an electronic interface to a freely moving air coil. Two delivery systems have been developed, one for peripheral joints and one for the axial skeleton. The most effective magnetic field configuration delivered by these devices was established empirically by treating groups of 20 patients with rheumatic conditions, varying the field variables until the optimal configuration was determined. Uncontrolled observations were then carried out in Europe in 861 patients with various painful rheumatic conditions; data suggested improvement of symptoms in 70–80% of treated patients. Therefore, a carefully controlled, double blind pilot study of possible therapeutic effects of this device was undertaken.

From the Department of Medicine (Rheumatology), Danbury Hospital, Danbury, and the Department of Surgery, St. Mary's Hospital, Waterbury, CT, USA.

Funded by Bio-Magnetic Therapy Systems, Inc.

D.H. Trock, MD, Assistant Chief of Rheumatology, Danbury Hospital, and Assistant Professor of Clinical Medicine, Yale University School of Medicine; A.J. Bollet, MD, Chief Section of Rheumatology, Danbury Hospital and Clinical Professor of Medicine, Yale University School of Medicine; R.H. Dyer, Jr, MD, Attending Orthopedist of St. Mary's and Waterbury Hospitals; L.P. Fielding, MD, Chairman, Department of Surgery, St. Mary's Hospital, and Professor of Surgery, Yale University School of Medicine; W.K. Miner, PA; R. Markoll, MD, PhD.

Address reprint requests to Dr. D.H. Trock, Section of Rheumatology, Danbury Hospital, Danbury, CT 06810.

Submitted January 7, 1992 revision accepted August 5, 1992.

MATERIALS AND METHODS

Patients. All patients met the criteria for the diagnosis of osteoarthritis (OA) published by Altman¹⁰. Radiographs were available in all but one of the patients with OA; severity grading was done by the criteria of Brandt¹¹.

We followed the approach to data gathering and analysis suggesting that the impact of a new drug or device on chronic arthritis is best determined by the patients' accounts of their joint pain and their ability to perform activities of daily living (ADL)^{12–14}.

Patients were required to be older than 18 years of age with persistent arthritic symptoms of at least one year duration, incompletely relieved by conventional treatment including nonsteroidal antiinflammatory drugs (NSAID), other analgesics, and physical therapy modalities. Patients who

had started any new form of treatment, including NSAID, within one month were also excluded from the investigation. The single most symptomatic peripheral joint was studied (knee in 21, the first carpometacarpal (MCP) or interphalangeal group of joints of the hand in 5 and posttraumatic OA of the ankle in 1.) Patients with OA of isolated MCP or single proximal interphalangeal (PIP) joints, or OA of the spine were not accepted for the study.

All women of child bearing age had to agree to use contraceptives. Other exclusions included the use of a cardiac pacemaker or the presence of any serious, unstable medical illness. Informed consent for entry into a double blind trial was obtained.

Patients who were taking stable daily doses of NSAID were instructed not to change their usual medications during the study period; the use of medications was checked by history at each evaluation point, but no pill counts were done.

Although some patients with other forms of arthritis were evaluated, data are presented here only for the patients who met the published criteria for the diagnosis of OA, since too few patients with other diagnoses were studied to allow conclusions to be drawn.

Treatments. Treatments were administered by the device described above, which produces an extremely low frequency (less than 30 Hz.), varying, pulsed electromagnetic field averaging 10-20 gauss of magnetic energy at a coil current of up to 2 amperes drawn from a power source of 120 volts AC. The pulse phase duration was 67 ms, including 15 micropulses with a pause duration of 0.1 s. The wave duration varied according to the frequency used. The patients rested the joint being treated on a pillow, encircled by the air coil which did not contact the skin. Treatments were given for 30 min; 3-5 sessions were given each week for a total of 18 treatments, the entire treatment period extending over about one month.

The magnetic therapy system used comprised 3 components: the magnetic field generator (MFG), the electronic interface and the air coils of varying dimensions as described above. The MFG produces a low voltage, DC current. The gating circuits employed CMOS logic integrated circuits. The electronic interface carried the current to the air coil, forming a flexible link between the fixed MFG and the air coil. The winding geometry employed in the air coil is formed with a specially designed jig to produce a pure DC homogeneous magnetic field.

The air coil produced a uniform homogeneous magnetic flux throughout the x, y and z axes. The joint under treatment was positioned eccentrically in the air coil, completely within the magnetic field flux. Since the device applied a pure magnetic field through the air coil, no heat was generated, nor did the patients feel any sensation during the treatment.

The magnetic field flux engulfed the joint area; it can penetrate the entire joint area since neither the skin nor other tissues present a barrier to magnetic energy. Since the field consisted of a uniform (homogeneous) flux density, the size of the body part (joint) treated did not affect the quality or quantity of magnetic energy delivered. Therefore, there was little or no variability of dosage to various joints or body parts treated.

Randomization. Upon entry to the study, patients were randomized to receive active PEMF or placebo using a table of 1,000 random digits. A coded master record sheet was kept by an office administrator, and a trained PEMF therapist activated the coil energy of the PEMF device and timed each treatment to 30 min. The ON/OFF control of the MFG was in the "ON" position with the associated red light for all of the patients in both the active and placebo groups during treatments. The PEMF device produces no noise or sensation and therefore the placebo therapy was applied by not energizing the air coil; all patients kept the affected joint in the air coil for the same 30 min period signalled by a laboratory timing clock placed on the generator. In this way, both the patient and the examining physician remained blinded as to whether active PEMF or placebo was being given.

Data collection. A case report form was prepared for each patient which recorded their enrollment eligibility criteria, pertinent clinical history, baseline medications, and examination. Baseline hematocrit and white blood cell counts (WBC), serum electrolytes, creatinine, and an erythrocyte

sedimentation rate (ESR), as well as a urinalysis, were done to exclude any concurrent illness and to monitor for any changes. These tests were repeated at the end of the treatment period. A radiograph of the joint to be treated was taken unless one had been obtained within the preceding 4 months.

Evaluations by the patient and the physician observers were made at 4 points during the study: baseline, midway through the treatment series, at the end of the treatment period, and 4 weeks after treatment was completed. At each evaluation the patient's opinion about joint symptoms was recorded along with the physician's assessment of the patient's symptoms and clinical signs. Data were collected for overall severity of pain, difficulty performing the activities of daily living (ADL) identified as the most troublesome by the patient before therapy was begun, pain generated by those specific activities of daily living, and the worst discomfort experienced by the patient in the previous week. A standard visual analog pain scale was marked by the patient to indicate the total amount of pain, pain with ADL and worst discomfort in the past week; the scale had no markings but was 100 mm long and the patient's mark was measured from the zero point. The difficulty in performing the specific ADL identified as most troublesome were quantitated on a scale of 1 to 5 using explanatory adjectives of "none, slight, moderate, severe, or extreme" to assist patients in their evaluations.

Data were also recorded for a rheumatologist's evaluation of pain on joint motion and tenderness of the joint being treated to firm palpation. At each observation after baseline, the physician made an "overall assessment of improvement" on a 5-point scale (worse or no change = 1, slight but insignificant change = 2, moderate improvement = 3, excellent improvement = 4, and complete disappearance of symptoms in affected joint = 5).

Statistical methods. Data were analyzed for change from baseline to each observation point for each patient by nonparametric matched pair 2-tailed t tests using the Wilcoxon signed rank test, for both treated and control groups. Instat (Graphpad Software Co.) was used for statistical calculations. The physicians' overall assessment of improvement was analyzed at the end of the treatment period and one month later, comparing treated and placebo groups, also using 2-tailed t tests.

RESULTS

Fifteen patients with OA were randomized into the active treatment group, 12 into the placebo treatment group. Two patients, after agreeing to randomization, withdrew from the study before the start of active treatment and asked for unblinded, active treatment; one (randomized to active treatment) withdrew because of transportation difficulties; one (randomized to the placebo group) was hospitalized because of a hernia; both of these patients had been evaluated at baseline, but not after treatment had begun and were not included in the data analysis. Five patients did not appear for the evaluation one month after completion of treatment; 4 had received active treatment, one, in the placebo group, had been hospitalized for community acquired pneumonia. All data for each patient were included in the analysis, including the last observation point; the numbers evaluated at each point are given in Table 1.

The 2 groups of patients did not differ significantly in respect to age, sex, race, body weight, or number of years with symptoms. The active treatment group included 11 with OA of the knee, 3 of the hand (PIP and MCP joints); 1 had OA of the ankle. The placebo group included 10 with OA of the knee, 2 of the hand. The dropouts all had OA of the knees.

At baseline there were no significant differences between